

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 4/18/2014

HHS
mmB

Instrument Serial Number: 010446

As of 18-Apr-14

Start Date	End Date	Agency Name
18-Apr-14		DFS Central Lab
15-Nov-12	18-Apr-14	Meherrin River Regional Jail
19-Dec-11	15-Nov-12	DFS Central Lab
09-Dec-11	19-Dec-11	Chesterfield County PD
08-Dec-11	09-Dec-11	DFS Central Lab
25-Jul-11	08-Dec-11	Chesterfield County PD
15-Oct-10	25-Jul-11	DFS Central Lab
09-Oct-09	15-Oct-10	Chesterfield County SO
30-Apr-09	09-Oct-09	DFS Central Lab
13-Nov-08	30-Apr-09	Nelson County Sheriff's Office
19-May-08	13-Nov-08	DFS Central Lab

HNS
mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010446 Worksheet Start Date 4/18/2014
Location Meherrin River Regional Jail
Address 9000 Boydton Plank Rd, Alberta, VA 23821
DFS Technician Heather Stanton License No. 18910

☐ Laboratory

☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 762 Reference Barometer (mm HG) 764
Reference Barometer(RB)Serial # 009111 RB Calibration Due 5/23/2014

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	
0.300	0.301	0.291	0.310	0.305	
Precision		sample min	sample max	Sample 2	
0		0.305	0.305	0.305	
				Sample 3	

Standard (sea level)	PA Target	minimum	maximum	Sample 1	
0.100	0.100	0.097	0.103	0.102	
Precision		sample min	sample max	Sample 2	
0		0.102	0.102	0.102	
				Sample 3	


Dry gas standard Lot No. (with tank no.) AG320502-06

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010446

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Stark

Date

4/18/14

Issuing Analyst

Michelle Bullen

Date

4/21/14msk




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

hms
mmb

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	---------------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 04/18/2014
TEST EQUIPMENT NUMBER 010446		

RESULTS: TIME SAMPLE TAKEN 12:03 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 18 Apr 2014 at 15:48

AMS
mmB

Test Results

Instrument Serial Number 010446

Test # 001953

Subject Test

Test Location 1 Department of
Test Date 18 Apr 2014

Test Location 2 Forensic Science
Test Time 11:56
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 18 Apr 2014 End Time 12:04

Result Time 12:03

Result Date 18 Apr 2014

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 11:56

Data Type BLK

Sample Value 0.000

Sample Time 11:57

Data Type CHK

Sample Value 0.102

Sample Time 11:58

Data Type BLK

Sample Value 0.000

Sample Time 11:59

Data Type SUBJ

Sample Value 0.000

Sample Time 12:00

Data Type BLK

Sample Value 0.000

Sample Time 12:01

Data Type SUBJ

Sample Value 0.000

Sample Time 12:03

Data Type BLK

Sample Value 0.000

Sample Time 12:04

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG320502-06

Standard Expiration Date 07/24/2015

Tank Pressure 1026

Barometric Pressure 762 mmHg

Blow Sample Number 1 Blow Duration 3.53 sec

Blow Volume 1804 cc End-of-Blow Time 12:00

Blow Sample Number 2 Blow Duration 3.34 sec

Blow Volume 1821 cc End-of-Blow Time 12:03

Tamper Evident Stamp bd311530

Test Status Code 0

Test Status Success

HMS
mmb

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010446 Test Number: 1954

Test Date: 04/18/2014 Test Time: 12:06 EDT

Dry Gas Target: 0.100

Lot Number: AG320502-06 Exp Date: 07/24/2015

Tank Pressure: 1022 psi Barometric Pressure: 762 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:06
CHK	0.102	12:07
BLK	0.000	12:09
CHK	0.102	12:10
BLK	0.000	12:11
CHK	0.102	12:12

Test Status: *Success*

Calibration CRC: 8312375D

HIS
mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010446 Test Number: 1955

Test Date: 04/18/2014 Test Time: 12:15 EDT

Dry Gas Target: 0.301

Lot Number: AG321804-03 Exp Date: 08/06/2015

Tank Pressure: 85 psi Barometric Pressure: 762 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:15
CHK	0.305	12:16
BLK	0.000	12:18
CHK	0.305	12:19
BLK	0.000	12:20
CHK	0.305	12:21

Test Status: *Success*

Calibration CRC: 8312375D